

PT REPORT

Number: RIAB79.48 - Final
Page 1 of 2

Description	RIAB - Ring Investigation Identification Asbestos in Bulk Materials.
Participant	CRB Analyse Service GmbH Bahnhofstrasse 14 37181 Hardegsen Germany
Method and evaluation	The PT was organised in accordance with document: Protocol RIAB. After obtaining the reference value, the samples have been shipped to the participant's laboratory. The results of the participant are compared with the reference values. The results of the comparison are presented on page 2.
Policy statement:	The evaluation reports of VSL's proficiency testing schemes are provided for the purpose of communicating the proficiency demonstrated by participants on specific measurements and/or calibrations. The reports are intended to be used in support of demonstrating competence in measurement or calibration, fulfilling quality control requirements as stipulated in written standards on showing such competence, and claims of calibration and measurement capabilities.

Delft, 13 November 2019
VSL B.V.

R.P. van Otterloo
Project manager comparisons


VSL
Dutch
Metrology
Institute

P T R E P O R T

Number: RIAB79.48 - Final
Page 2 of 2

Period of ILC August 2019 – November 2019

Results & Evaluation

Fibretype	7901		7902	
	Results	Reference	Results	Reference
Chrysotile	5-10%	2-5%	10-15%	10-15%
Amosite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Crocidolite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Tremolite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Actinolite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Anthophyllite	< 0,1%	< 0,1%	< 0,1%	< 0,1%

Points: 10

Points: 10

Fibretype	7903		7904	
	Results	Reference	Results	Reference
Chrysotile	10-15%	10-15%	< 0,1%	< 0,1%
Amosite	2-5%	2-5%	< 0,1%	< 0,1%
Crocidolite	2-5%	0,1 -2%	< 0,1%	< 0,1%
Tremolite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Actinolite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Anthophyllite	< 0,1%	< 0,1%	< 0,1%	< 0,1%

Points: 10


Points: 10

Results

The result of RIAB 79 with a total score of 100 is satisfactory

Result over RIAB 76-79

The result of RIAB 76 to 79 with a total score of 97 is satisfactory





RINGONDERZOEK IDENTIFICATIE ASBEST IN BULKMATERIALEN (RIAB)

Voorlopige resultaten van alle deelnemende laboratoria: ronde 79
Overzicht van lopende scores rondes 76-79

Labnummer	Score 79-1	Score 79-2	Score 79-3	Score 79-4	Score ronde 79 (%)	Score rondes 76-79 (%)
1	10	10	10	10	100	94
2	10	10	10	10	100	100
9	10	10	10	10	100	100
11	10	10	10	10	100	100
13	10	10	10	10	100	97
19	10	10	10	10	100	89
24	10	10	10	10	100	97
28	10	10	10	10	100	97
32	10	10	8	10	95	99
33	10	10	10	10	100	100
34	10	10	10	10	100	97
35	10	10	10	10	100	97
36	10	10	10	10	100	98
39	10	10	10	10	100	97
40	10	10	10	10	100	87
45	10	10	10	10	100	95
48	10	10	10	10	100	97
49A	10	7	8	10	88	84
49B	7	10	7	10	85	89
50	10	10	10	10	100	n/a
51	10	10	10	10	100	n/a
52	7	7	2	10	65	n/a

n/a=not available