

PT REPORT

Number: RIAB81.48 - Final

Page 1 of 2

Description	RIAB - Ring Investigation Identification Asbestos in Bulk Materials.
Participant	CRB Analyse Service GmbH Bahnhofstrasse 14 37181 Hardegsen Germany
Method and evaluation	The PT was organised in accordance with document: Protocol RIAB. After obtaining the reference value, the samples have been shipped to the participants laboratory. The results of the participant are compared with the reference values. The results of the comparison are presented on page 2.
Policy statement:	The evaluation reports of VSL's proficiency testing schemes are provided for the purpose of communicating the proficiency demonstrated by participants on specific measurements and/or calibrations. The reports are intended to be used in support of demonstrating competence in measurement or calibration, fulfilling quality control requirements as stipulated in written standards on showing such competence, and claims of calibration and measurement capabilities.

Delft, 20 April 2020
VSL B.V.



VSL
Dutch
Metrology
Institute

R.P. van Otterloo
Project manager comparisons

PT REPORT

Number: RIAB81.48 - Final

Page 2 of 2

Period of ILC February 2020 – April 2020


Results & Evaluation

Fibre type	8101		8102	
	Results	Reference	Results	Reference
Chrysotile	10-15%	10-15%	5-10%	2-5%
Amosite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Crocidolite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Tremolite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Actinolite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Anthophyllite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Points: 10		Points: 10		

Fibre type	8103		8104	
	Results	Reference	Results	Reference
Chrysotile	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Amosite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Crocidolite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Tremolite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Actinolite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Anthophyllite	2-5%	5-10%	< 0,1%	< 0,1%
Points: 10		Points: 10		

Results The result of RIAB 81 with a total score of 100 is satisfactory

Result over RIAB 78-81 The result of RIAB 78 to 81 with a total score of 97 is satisfactory



RIAB 81 results

