



CRB Analyse Service GmbH
Stefan Pierdzig
Bahnhofstrasse 14
D-37181 Hardegsen
Germany

VSL
Thijsseweg 11
2629 JA Delft
P.O. Box 654
2600 AR Delft
The Netherlands
T 015 269 15 00
F 015 261 29 71
E info@vsl.nl
I www.vsl.nl

Your reference

Our reference
RIAB 2020-82

Your reference

subject
RIAB 82 result

Dear participant,

Please find enclosed the evaluation report of RIAB 82.

If you have any questions, please don't hesitate to contact us.

Regards


VSL PT-Staff
Peter van Otterloo

Dutch
Metrology
Institute



Dutch
Metrology
Institute

PT REPORT

Number: RIAB82.48 - Final
Page 1 of 2

Description: RIAB - Ring Investigation Identification Asbestos in Bulk Materials.

Participant: CRB Analyse Service GmbH
Bahnhofstrasse 14
37181 Hardegsen
Germany

Method and evaluation:
The PT was organised in accordance with document: Protocol RIAB.
After obtaining the reference value, the samples have been shipped to the participants laboratory.
The reported results of the participant are compared with the reference values.
The results of the comparison are presented on page 2.

Policy statement:
The evaluation reports of VSL's proficiency testing schemes are provided for the purpose of communicating the proficiency demonstrated by participants on specific measurements and/or calibrations. The reports are intended to be used in support of demonstrating competence in measurement or calibration, fulfilling quality control requirements as stipulated in written standards on showing such competence, and claims of calibration and measurement capabilities.

Delft, 20 August 2020
VSL B.V.

R.P. van Otterloo
Project manager comparisons



Dutch
Metrology
Institute

P T R E P O R T

Number: RIAB82.48 - Final
Page 2 of 2

Period of ILC June 2020 – August 2020

Results &
Evaluation

Fibre type	8201		8202	
	Results	Reference	Results	Reference
Chrysotile	10-15%	5-10%	< 0,1%	< 0,1%
Amosite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Crocidolite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Tremolite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Actinolite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Anthophyllite	< 0,1%	< 0,1%	< 0,1%	< 0,1%

Points: 10

Points: 10

Fibre type	8203		8204	
	Results	Reference	Results	Reference
Chrysotile	30-60%	30-60%	10-15%	10-15%
Amosite	< 0,1%	< 0,1%	5-10%	15-30%
Crocidolite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Tremolite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Actinolite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Anthophyllite	< 0,1%	< 0,1%	< 0,1%	< 0,1%

Points: 10


Points: 8

Results

The result with a total score of 95 is satisfactory

Result over
RIAB 79-82

The result of RIAB 79 to 82 with a total score of 95 is satisfactory



VSL B.V.

Thijsseweg 11, 2629 JA Delft
P.O. Box 654, 2600 AR Delft, Netherlands
phone +31 15 2691500
fax +31 15 2612971
vsl@vsl.nl
www.vsl.nl

This report is issued under the provision that no liability is accepted and that the applicant gives warranty for each responsibility against third parties.

Reproduction of the complete report is permitted. Parts of this report may only be reproduced after written permission from the issuer.

