



CRB Analyse Service GmbH
Stefan Pierdzig
Bahnhofstrasse 14
D-37181 Hardegsen - Germany

VSL
Thijssseweg 11
2629 JA Delft
P.O. Box 654
2600 AR Delft
The Netherlands

T 015 269 15 00
F 015 261 29 71
E info@vsl.nl
I www.vsl.nl

Your reference

Our reference
RIAB 2021-85

Your reference

subject
RIAB 85 results

Dear participant,

Please find enclosed the evaluation report of RIAB 85.

If you have any questions, please don't hesitate to contact us.

Regards

VSL PT-Staff
Peter van Otterloo

Description RIAB - Ring Investigation Identification Asbestos in Bulk Materials.

Participant CRB Analyse Service GmbH
Bahnhofstrasse 14
37181 Hardegsen
Germany

Method and evaluation The PT was organised in accordance with document: Protocol RIAB.
After obtaining the reference value, the samples have been shipped to the participants laboratory.
The reported results of the participant are compared with the reference values.
The results of the comparison are presented on page 2.

Policy statement:

The evaluation reports of VSL's proficiency testing schemes are provided for the purpose of communicating the proficiency demonstrated by participants on specific measurements and/or calibrations. The reports are intended to be used in support of demonstrating competence in measurement or calibration, fulfilling quality control requirements as stipulated in written standards on showing such competence, and claims of calibration and measurement capabilities.

Delft, 16 April 2021
VSL B.V.

R.P. van Otterloo
Project manager comparisons



Dutch
Metrology
Institute

Period of ILC February 2021 – April 2021

Results &
Evaluation

Fibre type	8501		8502	
	Results	Reference	Results	Reference
Chrysotile	10-15 %	2-5 %	15-30 %	15-30 %
Amosite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Crocidolite	< 0.1%	< 0.1%	< 0.1%	0.1-2%
Tremolite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Actinolite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Anthophyllite	< 0,1%	< 0,1%	< 0.1%	< 0.1%
	Points: 8		Points: 7	

Fibre type	8503		8504	
	Results	Reference	Results	Reference
Chrysotile	0.1-2%	< 0.1%	30-60%	30-60%
Amosite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Crocidolite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Tremolite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Actinolite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Anthophyllite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
	Points: 0		Points: 10	

Results The result with a total score of 63 is questionable

RIAB 85 results

