

# PT REPORT


Number: RIAB87.48 - Final

Page 1 of 2

Description	RIAB - Ring Investigation Identification Asbestos in Bulk Materials.
Participant	CRB Analyse Service GmbH Bahnhofstrasse 14 37181 Hardegsen Germany
Method and evaluation	The PT was organised in accordance with document: Protocol RIAB. After obtaining the reference value, the samples have been shipped to the participants laboratory. The reported results of the participant are compared with the reference values. The results of the comparison are presented on page 2.
Policy statement:	The evaluation reports of VSL's proficiency testing schemes are provided for the purpose of communicating the proficiency demonstrated by participants on specific measurements and/or calibrations. The reports are intended to be used in support of demonstrating competence in measurement or calibration, fulfilling quality control requirements as stipulated in written standards on showing such competence, and claims of calibration and measurement capabilities.

Delft, 3 November 2021

VSL B.V.

The image shows a handwritten signature in blue ink over a blue rectangular stamp. The stamp contains the VSL logo, which is a circular pattern of dots, and the text 'VSL Dutch Metrology Institute' below it.

R.P. van Otterloo

## VSL B.V.

Thijssseweg 11, 2629 JA Delft  
P.O. Box 654, 2600 AR Delft, Netherlands  
phone +31 15 2691500  
fax +31 15 2612971  
vsl@vsl.nl  
www.vsl.nl

This report is issued under the provision that no liability is accepted and that the applicant gives warranty for each responsibility against third parties.

Reproduction of the complete report is permitted. Parts of this report may only be reproduced after written permission from the issuer.

# PT REPORT

Number: RIAB87.48 - Final

Page 2 of 2

Period of ILC September 2021 – October 2021

Results &  
Evaluation

Fibre type	8701		8702	
	Results	Reference	Results	Reference
Chrysotile	5-10%	2-5%	< 0.1%	< 0.1%
Amosite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Crocidolite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Tremolite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Actinolite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Anthophyllite	< 0,1%	< 0,1%	< 0.1%	< 0.1%
	Points: 10		Points: 10	

Fibre type	8703		8704	
	Results	Reference	Results	Reference
Chrysotile	30-60%	30-60%	5-10%	0.1-2%
Amosite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Crocidolite	< 0.1%	0.1-2%	< 0.1%	< 0.1%
Tremolite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Actinolite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Anthophyllite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
	Points: 7		Points: 8	

Results The result with a total score of 88 is satisfactory

RIAB 87 results

