

PT REPORT

Number: RIAB – 89 (Final)

Page 1 of 2

Description	RIAB - Ring Investigation Identification Asbestos in Bulk Materials.
Participant	CRB Analyse Service GmbH Bahnhofstrasse 14 D-37181 Hardegsen Germany
Method and evaluation	The PT was organised in accordance with document: Protocol RIAB. After obtaining the reference value, the samples have been shipped to the participants laboratory. The reported results of the participant are compared with the reference values. The results of the comparison are presented on page 2.
Policy statement:	The evaluation reports of VSL's proficiency testing schemes are provided for the purpose of communicating the proficiency demonstrated by participants on specific measurements and/or calibrations. The reports are intended to be used in support of demonstrating competence in measurement or calibration, fulfilling quality control requirements as stipulated in written standards on showing such competence, and claims of calibration and measurement capabilities.

Delft, 28 April 2022
VSL B.V.


J. W. van der Hout
Coordinator Proficiency Testing



Dutch
Metrology
Institute

VSL B.V.

Thijssseweg 11, 2629 JA Delft
P.O. Box 654, 2600 AR Delft, Netherlands
phone +31 15 2691500
fax +31 15 2612971
vsl@vsl.nl
www.vsl.nl

This report is issued under the provision that no liability is accepted and that the applicant gives warranty for each responsibility against third parties.

Reproduction of the complete report is permitted. Parts of this report may only be reproduced after written permission from the issuer.

PT REPORT

Number: RIAB – 89 (Final)

Page 2 of 2

Period of ILC March 2022

Results &
Evaluation

Fibre type	8901		8902	
	Results	Reference	Results	Reference
Chrysotile	10-15%	5-10%	< 0.1%	< 0.1%
Amosite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Crocidolite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Tremolite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Actinolite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Anthophyllite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
	Points: 10		Points: 10	

Fibre type	8903		8904	
	Results	Reference	Results	Reference
Chrysotile	2-5%	2-5%	15-30%	15-30%
Amosite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Crocidolite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Tremolite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Actinolite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
Anthophyllite	< 0.1%	< 0.1%	< 0.1%	< 0.1%
	Points: 10		Points: 10	

Results The total score of this round is 40 points, 100% and is satisfactory